

Accuracy of pO₂ – a 20 year overview of blood gas performance

Author Names: M.A. Thomas, S. Jones, G. Davies
WeQas, Cardiff and Vale University Health Board, Cardiff, UK

Introduction

Over the last 2 decades industry has provided us with an increasing number of smaller, portable analysers, utilising single or multi-use cartridges for Point of Care (POCT) Blood Gas analysis. Most EQA Providers distribute aqueous material for performance assessment of blood gas analysers, however annual studies conducted by WeQas using tonometered haemolysate have identified significant matrix issues with aqueous material especially for pO₂. The aim of this retrospective study was to assess the performance of Laboratory and POCT analysers for pO₂ using both fresh haemolysate and aqueous material.

Method

Tonometered blood haemolysate with oxygen saturation kinetics identical to that of fresh blood was distributed on an annual basis to all participants in the WeQas Blood Gas Scheme over a 20-year period. The performance for pO₂, expressed as coefficient of variation (CV) and bias against the All Method Average was calculated for each analyser for the haemolysate and concentration matched aqueous samples. The All Method Average was used in preference to the overall mean to allow equal weighting for each method group.

Review of Results 2010 to 2022

An increase in the use of POCT analysers was observed over this period, representing 95% of analysers in 2023, 43% in 2010 and 7% in 2001 respectively, Table 1. There was little change in the overall precision profile from 2001 to 2021 using the aqueous based material, Figure 1, however a much lower overall method CV was observed at a pO₂ < 15 kPa for analysers prior to 2000. At low pO₂ concentration, there was a significant improvement in the interlaboratory variation for the haemolysate material compared with aqueous material, Figure 2. A CV of 14% and 30% at a concentration of 8kPa, and a CV of 10% and > 50% at a concentration of 4kPa was observed for the haemolysate and aqueous materials respectively. The analyser interlaboratory variation expressed as ± 2 SD from the instrument mean and the bias to the Average of all method mean at a concentration of ≥ 4 kPa is shown in Figures 3 a to e over the 20 year period.

Figure 1. Precision profile for pO₂ using proteinated aqueous material from 1995 to 2020




Figure 2. Precision profile for pO₂ using fresh Tonometered haemolysate material from 2004 to 2023





Table 1. Change in Blood gas analyser type and growth of POCT analysers from 2001 to 2023

Analyser Type	Year					
	2023	2020	2016	2010	2006	2001
Conventional electrodes	127	244	541	947	853	505
Multi use Cartridge	1876	1440	1033	541	239	14
Optical sensor	13	16	28	39	34	13
Single use Cartridge	427	338	269	137	84	11
% POCT Analysers	94.8	88	71.1	43.1	29.5	7
Total	2443	2038	1871	1664	1210	543

Figure 3a,b,c,d,e. Performance of Blood Gas analysers for pO₂ over the last 2 decades assessed using Fresh tonometered haemolysate material

2004



2007

